Safe Use of Ceiling Hoist

OPERATING PROCEDURE

Model:
FIX-S
Maximum load:
200 kg (!)

HEALTHCARE USE / RECHARGEABLE BATTERY / MANDATORY SAFETY / PERIODIC MAINTENANCE

1 CHECKS BEFORE EACH USE

- Visually inspect the condition of the hoist strap, spreader bar, and sling straps.
- Visually inspect the end stops and the condition of the rail.
- Consult the patient risk assessment document (including sling selection and size).
- Follow the hoist and sling instructions.
- Verify that the hoist moves freely throughout the entire track.

2 OPERATION

LIFTING

- Lower the spreader bar using the control handset.
- Correctly attach the sling straps.
- Lift the patient smoothly and in a controlled way.

TRANSFERRING

- Move the patient safely.
- Do not pull the control handset to move the hoist.
- Maintain a stable position during movement.

Controls: ▲ Up ▼ Down

3 BATTERY CHARGING

- Place the hoist handset in its wall-mounted charging cradle.
- Verify that charging is in progress (orange light illuminated on the charger).
- Avoid completely discharging the battery.



4 MAINTENANCE AND CLEANING

- Use non-abrasive cleaning products.
- Do not wet the hoist. Clean with a damp cloth.
- Periodic maintenance required.
- Daily visual inspection.
- Annual technical inspection mandatory.
- Replace worn components only through an authorised service provider.

5 COMMON PROBLEMS AND SOLUTIONS

PROBLEM	ACTION
Not working	Check the battery and emergency stop
Low battery warning	Place on charge
Irregular movement	Stop use and notify maintenance
Will not lower	Use emergency lowering


Warnings

- Read these instructions completely before using the equipment.
- Do not exceed the maximum rated load.
- Do not use the hoist if the strap is damaged.
- Use only approved and suitable slings.
- A fully discharged battery may suffer permanent damage.

In case of emergency

Use only in the event of a failure of the main system.

1 EMERGENCY LOWERING



2 EMERGENCY STOP

Pull the **red cord.**

QUICK TRAINING

Scan the QR code

Carima
Solutions that move

Tel. +34 943 06 20 16
service@carimatech.com
carimatech.com/en